

**DDR 12 | RIGID TEST FINGER**  
IEC 61032 - Test Probe 11



**SECOM**

## DDR 12 | RIGID TEST FINGER

IEC 61032 - Test Probe 11



Precision tip



### Description of the device

The Rigid Test Finger is a precision test probe made according to Figure 7 (Fig. 7) of the **IEC 61032** (Test probe 11) and is used to simulate a human finger. It is also used by the standards of CSA, IRAM, UL, IEC 60335, IRAM 4220-1 and in most of the rules involved in the verification of accessibility to live parts.

The Rigid Test Finger was made on stainless steel and Polyamide handle.



### Technical characteristics

- Material: Stainless Steel.
- Handle: Polyamide.
- Surface finishing: Polished.
- According to: IEC 61032 - Test Probe 11.

SECOM reserves the right to change the specifications or design without prior notice.